

# Certificate of Compliance

**Certificate:** 1592458 (80259) **Master Contract:** 157918  
**Project:** 1671324 **Date Issued:** 2005/05/06  
**Issued to:** **BW Technologies Limited**  
2840 2nd Ave SE  
Calgary, AB T2A 7X9  
Canada  
Attention: Mr. Thom Crawford

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'*



**Issued by:** Wesley Van Hill

**Authorized by:** Patricia Pasemko, Operations Manager



## **PRODUCTS**

- CLASS 4828 82** - SIGNAL APPLIANCES-Toxic Gas Detection Instruments-For Hazardous Locations. Certified to U.S. Standards
- CLASS 4828 81** - SIGNAL APPLIANCES-Combustible Gas Detection Instruments - For Hazardous Location-Certified to U.S. Standards
- CLASS 4828 02** - SIGNAL APPLIANCES - Toxic Gas Detection Instruments - For Hazardous Locations
- CLASS 4828 01** - SIGNAL APPLIANCES - - Combustible Gas Detection Instruments - For Hazardous

Class I, Division 1, Groups A,B,C,D

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



**Certificate:** 1592458 (80259)

**Master Contract:** 157918

**Project:** 1671324

**Date Issued:** 2005/05/06

---

GasAlert Micro 5 Portable Gas Detector, Model M5-xwt1t2-r-p-d-a-b-cc & M5PID-xwt1t2-r-p-d-a-b-cc; utilizing electrochemical, catalytic bead and photo-ionization sensors; Intrinsically Safe when powered by one of the following AA Size Batteries / Battery Pack

- Duracell MN1500; T-Code T4; Ambient -20 to +40°C; T-Code 139.8°C (T3C); Ambient -20 to +50°C

- Energizer E91; T-Code 153°C(T3C); Ambient -20 to +40°C; T-Code 163°C (T3B); Ambient -20 to +50°C

- NiMH Rechargeable Battery Pack "M5-BAT01"; T-code T4; Ambient -20 to +50°C

#### **APPLICABLE REQUIREMENTS**

C22.2 No 0 - M1991 - General Requirements - Canadian Electrical Code Part II.

C22.2 No 152 - M1984 - Combustible Gas Detection Instruments.

C22.2 No 157 - M1992 - Intrinsically Safe and Non-Incendive Equipment for Use in Hazardous Locations.

UL 913, Sixth Edition - Intrinsically Safe Apparatus and Associated Apparatus for use in Class I, II, III, Division 1, Hazardous (Classified) Locations.

ISA S12.13.01-2000 - Performance Requirements, Combustible Gas Detectors



## *Supplement to Certificate of Compliance*

**Certificate:** 1592458

**Master Contract:** 157918

*The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
1671324	2005/05/06	Update to Report 1592458 to include alternate battery pack, minor construction changes and firmware update.
1592458	2005/03/31	cCSAus Certification of Portable Gas Detector for Class I, Division 1, Groups A, B, C and D